



財團法人全國認證基金會
Taiwan Accreditation Foundation

Certificate of Accreditation

(Certificate No: L0382-230823)

This is to certify that

Taiwan Testing and Certification Center Calibration Laboratory of Hsinchu

2F-5, No.47, Park Avenue II, Science-Based Industrial Park, Hsinchu 300, Taiwan (R.O.C.)

is accredited in respect of laboratory

Accreditation Criteria : ISO/IEC 17025: 2017; CNS 17025: 2018

Accreditation Number : 0382

Originally Accredited : June 15, 1999

Effective Period : August 09, 2023 to August 08, 2026

Accredited Scope : Calibration Field, see described in the Appendix



Scan to verify

Ching-Chang Lien

Ching-Chang Lien
President, Taiwan Accreditation Foundation
August 23, 2023

Accreditation Number : 0382

Laboratory Head : YEH, Ming-Hai

Mass/Force

| calibration items | working standard | calibration method | measurand level or range | | | | measurement conditions /independent variable | smallest uncertainty | |
|---|------------------|--------------------|--------------------------|-------|---------------|-------|--|----------------------|-------|
| | brand /model | document name /no. | minimum value | units | maximum value | units | explanation | value | units |
| KC4002 Torque Screwdriver Tester Nobar /NORBAR TST2 Nobar /NORBAR TST10 | | ISO 6789 | 0.02 | N·m | <0.04 | N·m | C.W. | 15 | % |
| | | | 0.04 | N·m | <0.2 | N·m | C.W. | 7.5 | % |
| | | | 0.2 | N·m | <0.4 | N·m | C.W. | 2.0 | % |
| | | | 0.4 | N·m | 4 | N·m | C.W. | 1.5 | % |
| | | | 0.02 | N·m | <0.04 | N·m | C.C.W. | 14 | % |
| | | | 0.04 | N·m | <0.2 | N·m | C.C.W. | 7.3 | % |
| | | | 0.2 | N·m | <0.4 | N·m | C.C.W. | 1.8 | % |
| | | | 0.4 | N·m | 4 | N·m | C.C.W. | 1.4 | % |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |



Pressure/Vacuum

| calibration items | working standard | calibration method | measurand level or range | | | | measurement conditions /independent variable | smallest uncertainty | |
|--|---|--------------------------|--------------------------|--------------------|---------------|------------|--|----------------------|------------|
| | | | brand /model | document name /no. | minimum value | units | | maximum value | units |
| KD1003 Mercury manometer /Atmospheric pressure gauge (On-site calibration included) | Pressure Indicator GE Druck /PACE1000 | Document No.: B00-CD-461 | 3.5 (35) | kPa (mbar) | 140 (1400) | kPa (mbar) | | 0.023 (0.23) | kPa (mbar) |
| | | | >140 (1400) | kPa (mbar) | 210 (2100) | kPa (mbar) | | 0.025 (0.25) | kPa (mbar) |
| | | | >210 (2100) | kPa (mbar) | 280 (2800) | kPa (mbar) | | 0.025 (0.25) | kPa (mbar) |
| | | | >280 (2800) | kPa (mbar) | 350 (3500) | kPa (mbar) | | 0.026 (0.26) | kPa (mbar) |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |
| KD1004 Pressure gauge (OIL PRESSURE GAUGE) (on-site calibration included) | Pressure gauge PDK /PDR1000 PDK /PDR500 | Document No.: B00-CD-550 | 0 | MPa | 150 | MPa | OIL PRESSURE GAUGE | 0.37 | MPa |
| | | | >150 | MPa | 250 | MPa | OIL PRESSURE GAUGE | 0.5 | MPa |
| | | | >250 | MPa | 500 | MPa | OIL PRESSURE GAUGE | 3.1 | MPa |
| | | | 0 | MPa | 50 | MPa | OIL PRESSURE GAUGE (On-site calibration) | 0.37 | MPa |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |



| calibration items | working standard | calibration method | measurand level or range | | | | measurement conditions /independent variable | smallest uncertainty | |
|--|--|-----------------------------|--------------------------|-------------------------------|--------------------|-------------------------------|--|----------------------|-------------------------------|
| | brand /model | document name /no. | minimum value | units | maximum value | units | explanation | value | units |
| KD1004 Pressure gauge (for Gas) (On-site calibration included) | Pressure gauge Additel/AD T681 Additel/AD T672 Fluke/PM-200-G4M Fluke/PM-200-G20M GE Druck/CM0-10G | Document No.: B00-CD-463 | -97 (-0.98) | kPa (kgf/cm ²) | -95 (-0.97) | kPa (kgf/cm ²) | AIR PRESSURE GAUGE | 0.064 (0.00065) | kPa (kgf/cm ²) |
| | | | >-95 (-0.97) | kPa (kgf/cm ²) | 0 (0) | kPa (kgf/cm ²) | AIR PRESSURE GAUGE | 0.023 (0.00023) | kPa (kgf/cm ²) |
| | | | >0 (0) | kPa (kgf/cm ²) | 700 (7.14) | kPa (kgf/cm ²) | AIR PRESSURE GAUGE | 0.12 (0.0012) | kPa (kgf/cm ²) |
| | | | >700 (7.14) | kPa (kgf/cm ²) | 2068.0 (21.09) | kPa (kgf/cm ²) | AIR PRESSURE GAUGE | 1.0 (0.010) | kPa (kgf/cm ²) |
| | | | >2068.0 (21.09) | kPa (kgf/cm ²) | 19995.5 (203.9) | kPa (kgf/cm ²) | AIR PRESSURE GAUGE | 9.4 (0.096) | kPa (kgf/cm ²) |
| | | | >19995.5 (203.9) | kPa (kgf/cm ²) | 20685 (210.9) | kPa (kgf/cm ²) | AIR PRESSURE GAUGE | 32 (0.33) | kPa (kgf/cm ²) |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |



| calibration items | working standard | calibration method | measurand level or range | | | | measurement conditions /independent variable | smallest uncertainty | |
|--|--|--------------------------|--------------------------|-------|---------------|-------|--|----------------------|-------|
| | brand /model | document name /no. | minimum value | units | maximum value | units | explanation | value | units |
| KD1007 Differential Pressure Gauge (On-site calibration included) | Pressure Indicator Pressure Calibrator FLUKE /RPM4 BG15Ks | Document No.: B00-CD-482 | -15 | kPa | 15 | kPa | | 1.7 | Pa |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |

Temperature/Humidity

| calibration items | working standard | calibration method | measurand level or range | | | | measurement conditions /independent variable | smallest uncertainty | |
|--|---|---|--------------------------|-------|---------------|-------|--|----------------------|-------|
| | brand /model | document name /no. | minimum value | units | maximum value | units | explanation | value | units |
| KE1002 Platinum Resistance Thermometer | Platinum Resistance Thermometer Fluke/5628 Fluke/1529 Fluke/5609 Isotech/Venus 2140 /935-14-82 (on-site calibration) | Document No.: B00-CD-474 Document No.: B00-CD-570 (on-site calibration) | -196 | °C | -196 | °C | | 0.058 | °C |
| | | | -193 | °C | -80 | °C | | 0.083 | °C |
| | | | >-80 | °C | 0 | °C | | 0.063 | °C |
| | | | >0 | °C | 130 | °C | | 0.058 | °C |
| | | | >130 | °C | 600 | °C | | 0.065 | °C |
| | | | -25 | °C | <0 | °C | on-site calibration | 0.3 | °C |
| | | | 0 | °C | 100 | °C | on-site calibration | 0.23 | °C |
| | | | >100 | °C | 135 | °C | on-site calibration | 0.3 | °C |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |



| calibration items | working standard | calibration method | measurand level or range | | | | measurement conditions /independent variable | smallest uncertainty | |
|---|--|--|--------------------------|-------|---------------|-------|--|----------------------|-------|
| | brand /model | document name /no. | minimum value | units | maximum value | units | explanation | value | units |
| KE1005 Thermocouple Thermometer (on-site calibration included) | Platinum Resistance | Document No.: B00-CD-475 | -193 | °C | -80 | °C | Type J, K, T | 0.59 | °C |
| | Thermometer | | -80 | °C | 400 | °C | Type T | 0.58 | °C |
| | Fluke/5628 | Document No.: B00-CD-571 (on-site calibration) | -80 | °C | 600 | °C | Type J, K | 0.58 | °C |
| | Fluke/1529 | | -25 | °C | 0 | °C | on-site calibration: TypeK, J, T | 0.7 | °C |
| | Fluke/5609 | | 0 | °C | 100 | °C | on-site calibration: TypeK, J, T | 0.63 | °C |
| | Isotech/Venus 2140/935-14-82 (on-site calibration) | | 100 | °C | 135 | °C | on-site calibration: TypeK, J, T | 0.7 | °C |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |
| KE1006 Temperature Meter (on-site calibration included) Temperature Recorder (on-site calibration included) | RESISTANCE DECADE BOX /IET HARS-X-8- 0.01 Multifunction Process Calibrator /Additel 222A | Document No.: B00-CD-568 | -200 | °C | 850 | °C | PT100 | 0.064 | °C |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |



| calibration items | working standard | calibration method | measurand level or range | | | | measurement conditions /independent variable | smallest uncertainty | |
|--|---|-----------------------------|--------------------------|--------|---------------|-------|--|----------------------|-------|
| | brand /model | document name /no. | minimum value | units | maximum value | units | explanation | value | units |
| KE1006 Temperature Meter (on-site calibration included) Temperature Recorder (on-site calibration included) | Thermocouple Simulator Eurotron /MicroCal 2000+ | Document No.: B00-CD-403 | -200 | °C | 800 | °C | J type | 0.52 | °C |
| | | | -200 | °C | 1000 | °C | K type | 0.48 | °C |
| | | | -200 | °C | 400 | °C | T type | 0.54 | °C |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |
| KE1008 Thermocouple Simulator (on-site calibration included) Temperature Calibrator (on-site calibration included) | Thermocouple Simulator Eurotron/MicroCal 2000+ | Document No.: B00-CD-403 | -200 | °C | -100 | °C | J type | 0.5 | °C |
| | | | -100 | °C | 800 | °C | J type | 0.52 | °C |
| | | | -200 | °C | -100 | °C | K type | 0.7 | °C |
| | | | -100 | °C | 1000 | °C | K type | 0.67 | °C |
| | | | -200 | °C | -100 | °C | T type | 0.8 | °C |
| -100 | °C | 400 | °C | T type | 0.66 | °C | | | |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |
| KE1009 Temperature Calibrator (on-site calibration included) | RESISTANCE DECADE BOX /IET HARS-X-8-0.01 Multifunction Meter /PicoTest M3522A | Document No.: B00-CD-568 | -200 | °C | 850 | °C | PT100 | 0.013 | °C |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |



| calibration items | working standard | calibration method | measurand level or range | | | | measurement conditions /independent variable | smallest uncertainty | |
|--|---|-----------------------------|--------------------------|-------|---------------|-------|--|----------------------|-------|
| | brand /model | document name /no. | minimum value | units | maximum value | units | explanation | value | units |
| KE1010 Temperature Test Chamber (on-site calibration included) | Recorder YOKOGAWA /MV2020+T-TYPE | Document No.: B00-CD-448 | -75 | °C | 250 | °C | temperature | 0.96 | °C |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |
| KE2001 Hygrometer (on-site calibration included) | Hygrometer ROTRONIC/HP22A Recorder YOKOGAWA/DX23 0-1-1 +PT100 | Document No.: B00-CD-295 | 10 | °C | 70 | °C | temperature | 0.5 | °C |
| | | | 30 | % | 95 | % | relative humidity (@ 10°C to 70°C) | 2.2 | % |
| | | | -60 | °C | 150 | °C | temperature | 0.66 | °C |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |
| KE2004 Temperature & Humidity Recorder (on-site calibration included) | Hygrometer ROTRONIC/HP22A Recorder YOKOGAWA/DX23 0-1-1 +PT100 | Document No.: B00-CD-295 | 10 | °C | 70 | °C | temperature | 0.5 | °C |
| | | | 30 | % | 95 | % | relative humidity (@ 10°C to 70°C) | 2.2 | % |
| | | | -60 | °C | 150 | °C | temperature | 0.66 | °C |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |
| KE2005 Temperature & Humidity Test Chamber (on-site calibration included) | Recorder YOKOGAWA/MV20 20+T-TYPE | Document No.: B00-CD-449 | 20 | °C | 95 | °C | temperature | 0.96 | °C |
| | | | 30 | % | 98 | % | relative humidity (@ 20°C to 95°C) | 3.1 | % |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |



| calibration items | working standard | calibration method | measurand level or range | | | | measurement conditions /independent variable | smallest uncertainty | |
|--|----------------------------------|-----------------------------|--------------------------|-----------|---------------|-----------|---|----------------------|-----------|
| | brand /model | document name /no. | minimum value | units | maximum value | units | explanation | value | units |
| KE2099 Water Activity Meter (relative humidity type) | Hygrometer ROTRONIC /HP22A | Document No.: B00-CD-295 | 30 (0.3) | % (Aw) | 95 (0.95) | % (Aw) | Relative humidity [Water Activity (Aw)]: 100 %RH=1.000 Aw | 2.2 (0.022) | % (Aw) |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |

Electricity

| calibration items | working standard | calibration method | measurand level or range | | | | measurement conditions /independent variable | smallest uncertainty | |
|---|--|-----------------------------|--------------------------|-------|---------------|-------|--|----------------------|-------|
| | brand /model | document name /no. | minimum value | units | maximum value | units | explanation | value | units |
| KF1001 DC Voltage Meter DC Voltage Source Electrometer (Includes the negative voltage Calibration, On Site Calibration) | Electrometer /Keithley /6517B Multimeter /Fluke/8508A Multifunction Calibrator /Fluke/5700A | Document No.: B00-CD-496 | 1 | V | 1 | V | DC Voltage Meter, Electrometer (@2 V) | 0.15 | mV/V |
| | | | 1.9 | V | 1.9 | V | DC Voltage Meter, Electrometer (@2 V) | 0.15 | mV/V |
| | | | 10 | V | 10 | V | DC Voltage Meter, Electrometer (@20 V) | 0.068 | mV/V |
| | | | 19 | V | 19 | V | DC Voltage Meter, Electrometer (@20 V) | 0.068 | mV/V |
| | | | 100 | V | 100 | V | DC Voltage Meter, Electrometer (@200 V) | 0.064 | mV/V |
| | | | 190 | V | 190 | V | DC Voltage Meter, Electrometer (@200 V) | 0.064 | mV/V |
| | | | 1000 | V | 1000 | V | DC Voltage Meter (@1000 V) | 0.06 | mV/V |
| | | | 0 | V | 0 | V | DC Voltage Meter, Electrometer (@100 V) | 0.28 | mV |
| | | | 0 | V | 0 | V | DC Voltage Meter (@1000 V) | 1.3 | mV |
| | | | 1 | V | 1 | V | DC Voltage Source, Electrometer (@2 V) | 0.15 | mV/V |
| | | | 1.9 | V | 1.9 | V | DC Voltage Source, Electrometer (@2 V) | 0.15 | mV/V |



| calibration items | working standard | calibration method | measurand level or range | | | | measurement conditions /independent variable | smallest uncertainty | |
|---|---|--------------------------|--------------------------|-------|---------------|-------|--|----------------------|-------|
| | brand /model | document name /no. | minimum value | units | maximum value | units | explanation | value | units |
| KF1001 DC Voltage Meter DC Voltage Source Electrometer (Includes the negative voltage Calibration, On Site Calibration) | Electrometer /Keithley /6517B Multimeter /Fluke/8508A Multifunction Calibrator /Fluke/5700A | Document No.: B00-CD-496 | 10 | V | 10 | V | DC Voltage Source, Electrometer (@20 V) | 0.068 | mV/V |
| | | | 19 | V | 19 | V | DC Voltage Source, Electrometer (@20 V) | 0.068 | mV/V |
| | | | 100 | V | 100 | V | DC Voltage Source, Electrometer (@200 V) | 0.064 | mV/V |
| | | | 190 | V | 190 | V | DC Voltage Source, Electrometer (@200 V) | 0.064 | mV/V |
| | | | 100 | V | 100 | V | DC Voltage Source, Electrometer (@100 V) | 0.036 | mV/V |
| | | | 1000 | V | 1000 | V | DC Voltage Source, Electrometer (@1000 V) | 0.06 | mV/V |
| | | | 0 | V | 0 | V | DC Voltage Source, Electrometer (@100 V) | 0.28 | mV |
| | | | 0 | V | 0 | V | DC Voltage Source, Electrometer (@1000 V) | 1.3 | mV |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |
| KF1001 DCV Source DC Volt Meter (on-site calibration included) | Multimeter Fluke/8508A Calibrator Fluke/5700A | Document No.: B00-CD-213 | 1 | mV | 10 | mV | (@200 mV) | 0.96 | mV/V |
| | | | 10 | mV | 199 | mV | (@200 mV) | 0.13 | mV/V |
| | | | 0.1 | V | 1.99 | V | (@2 V) | 16 | μV/V |
| | | | 1 | V | 19.9 | V | (@20 V) | 11 | μV/V |
| | | | 10 | V | 199 | V | (@200 V) | 30 | μV/V |
| | | | 200 | V | 1000 | V | (@1000 V) | 13 | μV/V |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |



| calibration items | working standard | calibration method | measurand level or range | | | | measurement conditions /independent variable | smallest uncertainty | |
|--|--|-----------------------------|--------------------------|----------|---------------|-------|--|----------------------|-------|
| | brand /model | document name /no. | minimum value | units | maximum value | units | explanation | value | units |
| KF1002 DCA Source DC Ampere Meter (on-site calibration included) | Multimeter Fluke/8508A Calibrator Fluke/5700A | Document No.: B00-CD-211 | 1 | μA | 190 | μA | (@200 μA) | 0.59 | mA/A |
| | | | 0.1 | mA | 19 | mA | (@2 mA) | 0.048 | mA/A |
| | | | 1 | mA | 19 | mA | (@20 mA) | 0.064 | mA/A |
| | | | 10 | mA | 190 | mA | (@200 mA) | 0.18 | mA/A |
| | | | 0.1 | A | 2 | A | (@2 A) | 0.23 | mA/A |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |
| KF1002 DC Current Source DC Current Meter Electrometer (Includes the negative Current Calibration, On-Site Calibration Included) | Electrometer Keithley/6517B Current Source Keithley/220 Calibrator Keithley/263 | Document No.: B00-CD-496 | 10 | pA | 10 | pA | (@20 pA) | 2.6 | mA/A |
| | | | 19 | pA | 19 | pA | (@20 pA) | 2.6 | mA/A |
| | | | 100 | pA | 100 | pA | (@200 pA) | 1.3 | mA/A |
| | | | 190 | pA | 190 | pA | (@200 pA) | 1.3 | mA/A |
| | | | 1 | nA | 1 | nA | (@2 nA) | 0.42 | mA/A |
| | | | 1.9 | nA | 1.9 | nA | (@2 nA) | 0.42 | mA/A |
| | | | 10 | nA | 10 | nA | (@20 nA) | 0.45 | mA/A |
| | | | 19 | nA | 19 | nA | (@20 nA) | 0.45 | mA/A |
| | | | 100 | nA | 100 | nA | (@200 nA) | 0.43 | mA/A |
| | | | 190 | nA | 190 | nA | (@200 nA) | 0.43 | mA/A |
| | | | 1 | μA | 1 | μA | (@2 μA) | 0.74 | mA/A |
| | | | 1.9 | μA | 1.9 | μA | (@2 μA) | 0.74 | mA/A |
| | | | 10 | μA | 10 | μA | (@20 μA) | 0.90 | mA/A |
| | | | 19 | μA | 19 | μA | (@20 μA) | 0.90 | mA/A |
| | | | 100 | μA | 100 | μA | (@200 μA) | 0.15 | mA/A |
| | | | 190 | μA | 190 | μA | (@200 μA) | 0.15 | mA/A |
| | | | 1 | mA | 1 | mA | (@2 mA) | 0.11 | mA/A |
| 1.9 | mA | 1.9 | mA | (@2 mA) | 0.11 | mA/A | | | |
| 10 | mA | 10 | mA | (@20 mA) | 0.10 | mA/A | | | |



| calibration items | working standard | calibration method | measurand level or range | | | | measurement conditions /independent variable | smallest uncertainty | |
|--|--|-----------------------------|--------------------------|-------|---------------|-------|--|----------------------|-------|
| | brand /model | document name /no. | minimum value | units | maximum value | units | explanation | value | units |
| KF1002 DC Current Source DC Current Meter Electrometer (Includes the negative Current Calibration, On-Site Calibration Included) | Electrometer Keithley/6517B Current Source Keithley/220 Calibrator Keithley/263 | Document No.: B00-CD-496 | 19 | mA | 19 | mA | (@20 mA) | 0.10 | mA/A |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |
| KF1003 High Digital Voltage Meter (on-site calibration included) High Digital Voltage Source (on-site calibration included) | High Digital Voltage Meter KIKUSUI 149- 10A | Document No.: B00-CD-406 | 1 | kV | 10 | kV | | 7.0 | mV/V |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |
| KF1004 DC Current Meter DC Current Source DC Current Shunt (Includes the negative Current Calibration, On Site Calibration) | Calibrator/Current Amplifier Fluke/5700A Fluke/52120A | Document No.: B00-CD-396 | 2 | A | 100 | A | Current Meter | 0.10 | mA/A |
| | | | 2 | A | 100 | A | Current Source | 0.14 | mA/A |
| | | | 2 | A | 100 | A | Current Shunt | 0.28 | mA/A |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |



| calibration items | working standard | calibration method | measurand level or range | | | | measurement conditions /independent variable | smallest uncertainty | |
|--|--|-----------------------------|--------------------------|--------------------|---------------|-------|--|----------------------|-------|
| | brand /model | document name /no. | minimum value | units | maximum value | units | explanation | value | units |
| KF1004 DC Clamp on Meter DC Current Meter DC Current Source (Includes the negative Current Calibration, On Site Calibration) | Current Amplifier Fluke/52120A Current Coil Fluke/52120A /Coil 3kA | Document No.: B00-CD-465 | 0 | A | 0 | A | | 0.77 | mA |
| | | | 0 | A | 2 | A | | 0.097 | mA/A |
| | | | 2 | A | 20 | A | | 0.088 | mA/A |
| | | | 20 | A | 100 | A | | 0.10 | mA/A |
| | | | 100 | A | 200 | A | | 0.24 | mA/A |
| | | | 200 | A | 1000 | A | | 0.28 | mA/A |
| | | | 1000 | A | 2500 | A | | 0.48 | mA/A |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |
| KF1011 ACV Source ACV Meter (on-site calibration included) | Multimeter Fluke/8508A Calibrator Fluke/5700A Keysight/33622A | Document No.: B00-CD-214 | 10 | mV | 10 | mV | ACV Source (@1 Hz) | 1.3 | mV/V |
| | | | 20 | mV | 20 | mV | ACV Source (@1 Hz) | 1.3 | mV/V |
| | | | 50 | mV | 50 | mV | ACV Source (@1 Hz) | 1.3 | mV/V |
| | | | 100 | mV | 100 | mV | ACV Source (@1 Hz) | 1.3 | mV/V |
| | | | 200 | mV | 200 | mV | ACV Source (@1 Hz) | 0.65 | mV/V |
| | | | 500 | mV | 500 | mV | ACV Source (@1 Hz) | 0.51 | mV/V |
| | | | 1 | V | 1 | V | ACV Source (@1 Hz) | 1.3 | mV/V |
| | | | 2 | V | 2 | V | ACV Source (@1 Hz) | 1.3 | mV/V |
| | | | 5 | V | 5 | V | ACV Source (@1 Hz) | 0.52 | mV/V |
| | | | 7 | V | 7 | V | ACV Source (@1 Hz) | 0.55 | mV/V |
| | | | 10 | mV | 10 | mV | ACV Source (@2 Hz) | 1.3 | mV/V |
| | | | 20 | mV | 20 | mV | ACV Source (@2 Hz) | 1.3 | mV/V |
| | | | 50 | mV | 50 | mV | ACV Source (@2 Hz) | 1.3 | mV/V |
| | | | 100 | mV | 100 | mV | ACV Source (@2 Hz) | 1.3 | mV/V |
| | | | 200 | mV | 200 | mV | ACV Source (@2 Hz) | 0.64 | mV/V |
| 500 | mV | 500 | mV | ACV Source (@2 Hz) | 0.51 | mV/V | | | |
| | | | 1 | V | 1 | V | ACV Source (@2 Hz) | 1.3 | mV/V |



| calibration items | working standard | calibration method | measurand level or range | | | | measurement conditions /independent variable | smallest uncertainty | |
|---|---|-----------------------------|--------------------------|---------------------|---------------|-------|--|----------------------|-------|
| | brand /model | document name /no. | minimum value | units | maximum value | units | explanation | value | units |
| KF1011 ACV Source ACV Meter (on-site calibration included) | Multimeter Fluke/8508A Calibrator Fluke/5700A Keysight/33622A | Document No.: B00-CD-214 | 2 | V | 2 | V | ACV Source (@2 Hz) | 1.3 | mV/V |
| | | | 5 | V | 5 | V | ACV Source (@2 Hz) | 0.51 | mV/V |
| | | | 7 | V | 7 | V | ACV Source (@2 Hz) | 0.55 | mV/V |
| | | | 10 | mV | 10 | mV | ACV Source (@5 Hz) | 1.3 | mV/V |
| | | | 20 | mV | 20 | mV | ACV Source (@5 Hz) | 1.3 | mV/V |
| | | | 50 | mV | 50 | mV | ACV Source (@5 Hz) | 1.3 | mV/V |
| | | | 100 | mV | 100 | mV | ACV Source (@5 Hz) | 1.3 | mV/V |
| | | | 200 | mV | 200 | mV | ACV Source (@5 Hz) | 0.65 | mV/V |
| | | | 500 | mV | 500 | mV | ACV Source (@5 Hz) | 0.51 | mV/V |
| | | | 1 | V | 1 | V | ACV Source (@5 Hz) | 1.3 | mV/V |
| | | | 2 | V | 2 | V | ACV Source (@5 Hz) | 1.3 | mV/V |
| | | | 5 | V | 5 | V | ACV Source (@5 Hz) | 0.51 | mV/V |
| | | | 7 | V | 7 | V | ACV Source (@5 Hz) | 0.55 | mV/V |
| | | | 10 | mV | 10 | mV | ACV Source (@10 Hz) | 1.3 | mV/V |
| | | | 20 | mV | 20 | mV | ACV Source (@10 Hz) | 1.3 | mV/V |
| | | | 50 | mV | 50 | mV | ACV Source (@10 Hz) | 1.3 | mV/V |
| | | | 100 | mV | 100 | mV | ACV Source (@10 Hz) | 1.3 | mV/V |
| | | | 200 | mV | 200 | mV | ACV Source (@10 Hz) | 0.65 | mV/V |
| | | | 500 | mV | 500 | mV | ACV Source (@10 Hz) | 0.51 | mV/V |
| | | | 1 | V | 1 | V | ACV Source (@10 Hz) | 1.3 | mV/V |
| 2 | V | 2 | V | ACV Source (@10 Hz) | 1.3 | mV/V | | | |
| 5 | V | 5 | V | ACV Source (@10 Hz) | 0.51 | mV/V | | | |
| 7 | V | 7 | V | ACV Source (@10 Hz) | 0.55 | mV/V | | | |
| 10 | mV | 10 | mV | ACV Source (@20 Hz) | 1.3 | mV/V | | | |
| 20 | mV | 20 | mV | ACV Source (@20 Hz) | 1.3 | mV/V | | | |
| 50 | mV | 50 | mV | ACV Source (@20 Hz) | 1.3 | mV/V | | | |



| calibration items | working standard | calibration method | measurand level or range | | | | measurement conditions /independent variable | smallest uncertainty | |
|---|---|-----------------------------|--------------------------|----------------------------------|---------------|-------|--|----------------------|-------|
| | brand /model | document name /no. | minimum value | units | maximum value | units | explanation | value | units |
| KF1011 ACV Source ACV Meter (on-site calibration included) | Multimeter Fluke/8508A Calibrator Fluke/5700A Keysight /33622A | Document No.: B00-CD-214 | 100 | mV | 100 | mV | ACV Source (@20 Hz) | 1.3 | mV/V |
| | | | 200 | mV | 200 | mV | ACV Source (@20 Hz) | 0.64 | mV/V |
| | | | 500 | mV | 500 | mV | ACV Source (@20 Hz) | 0.51 | mV/V |
| | | | 1 | V | 1 | V | ACV Source (@20 Hz) | 1.3 | mV/V |
| | | | 2 | V | 2 | V | ACV Source (@20 Hz) | 1.3 | mV/V |
| | | | 5 | V | 5 | V | ACV Source (@20 Hz) | 0.51 | mV/V |
| | | | 7 | V | 7 | V | ACV Source (@20 Hz) | 0.55 | mV/V |
| | | | 10 | mV | 199 | mV | ACV Source (Rnage: 200 mV@20 Hz) | 0.80 | mV/V |
| | | | 0.1 | V | 1.99 | V | ACV Source (Rnage: 2 V@20 Hz) | 0.24 | mV/V |
| | | | 1 | V | 19.9 | V | ACV Source (Rnage: 20 V@20 Hz) | 0.27 | mV/V |
| | | | 10 | mV | 199 | mV | ACV Source (Rnage: 200 mV@100 Hz) | 0.64 | mV/V |
| | | | 0.1 | V | 1.99 | V | ACV Source (Rnage: 2 V@100 Hz) | 0.10 | mV/V |
| | | | 1 | V | 19.9 | V | ACV Source (Rnage: 20 V@100 Hz) | 0.065 | mV/V |
| | | | 10 | mV | 199 | mV | ACV Source (Rnage: 20 V@3 kHz) | 0.14 | mV/V |
| | | | 0.1 | V | 1.99 | V | ACV Source (Rnage: 20 V@3 kHz) | 0.10 | mV/V |
| | | | 1 | V | 19.9 | V | ACV Source (Rnage: 20 V@3 kHz) | 0.10 | mV/V |
| | | | 10 | mV | 199 | mV | ACV Source (Rnage: 20 V@5 kHz) | 0.66 | mV/V |
| | | | 0.1 | V | 1.99 | V | ACV Source (Rnage: 20 V@5 kHz) | 0.12 | mV/V |
| | | | 1 | V | 19.9 | V | ACV Source (Rnage: 20 V@5 kHz) | 0.11 | mV/V |
| | | | 10 | mV | 199 | mV | ACV Source (Rnage: 20 V@10 kHz) | 0.64 | mV/V |
| 0.1 | V | 1.99 | V | ACV Source (Rnage: 20 V@10 kHz) | 0.12 | mV/V | | | |
| 1 | V | 19.9 | V | ACV Source (Rnage: 20 V@10 kHz) | 0.11 | mV/V | | | |
| 1 | mV | 10 | mV | ACV Source (Rnage: 200 mV@50 Hz) | 2.0 | mV/V | | | |
| 10 | mV | 199 | mV | ACV Source (Rnage: 200 mV@50 Hz) | 0.42 | mV/V | | | |
| 0.1 | V | 1.9 | V | ACV Source (Rnage: 2 V@50 Hz) | 0.13 | mV/V | | | |
| 1 | V | 19.9 | V | ACV Source (Rnage: 20 V@50 Hz) | 0.12 | mV/V | | | |



| calibration items | working standard | calibration method | measurand level or range | | | | measurement conditions /independent variable | smallest uncertainty | |
|---|---|-----------------------------|--------------------------|-------|---------------|-------|--|----------------------|-------|
| | brand /model | document name /no. | minimum value | units | maximum value | units | explanation | value | units |
| KF1011 ACV Source ACV Meter (on-site calibration included) | Multimeter Fluke/8508A Calibrator Fluke/5700A Keysight /33622A | Document No.: B00-CD-214 | 10 | V | 199 | V | ACV Source (Rnage: 200 V@50 Hz) | 0.12 | mV/V |
| | | | 100 | V | 1000 | V | ACV Source (Rnage: 1000 V@50 Hz) | 0.16 | mV/V |
| | | | 1 | mV | 10 | mV | ACV Source (Rnage: 200 mV@60 Hz) | 2.2 | mV/V |
| | | | 10 | mV | 199 | mV | ACV Source (Rnage: 200 mV@60 Hz) | 0.4 | mV/V |
| | | | 0.1 | V | 1.9 | V | ACV Source (Rnage: 2 V@60 Hz) | 0.13 | mV/V |
| | | | 1 | V | 19.9 | V | ACV Source (Rnage: 20 V@60 Hz) | 0.12 | mV/V |
| | | | 10 | V | 199 | V | ACV Source (Rnage: 200 V@60 Hz) | 0.12 | mV/V |
| | | | 100 | V | 1000 | V | ACV Source (Rnage: 1000 V@60 Hz) | 0.15 | mV/V |
| | | | 1 | mV | 10 | mV | ACV Source (Rnage: 200 mV@400 Hz) | 2.1 | mV/V |
| | | | 10 | mV | 199 | mV | ACV Source (Rnage: 200 mV@400 Hz) | 0.49 | mV/V |
| | | | 0.1 | V | 1.9 | V | ACV Source (Rnage: 2 V@400 Hz) | 0.11 | mV/V |
| | | | 1 | V | 19.9 | V | ACV Source (Rnage: 20 V@400 Hz) | 0.10 | mV/V |
| | | | 10 | V | 199 | V | ACV Source (Rnage: 200 V@400 Hz) | 0.10 | mV/V |
| | | | 100 | V | 1000 | V | ACV Source (Rnage: 1000 V@400 Hz) | 0.15 | mV/V |
| | | | 1 | mV | 10 | mV | ACV Source (Rnage: 200 mV@1 kHz) | 2.1 | mV/V |
| | | | 10 | mV | 199 | mV | ACV Source (Rnage: 200 mV@1 kHz) | 0.47 | mV/V |
| | | | 0.1 | V | 1.9 | V | ACV Source (Rnage: 2 V@1 kHz) | 0.11 | mV/V |
| | | | 1 | V | 19.9 | V | ACV Source (Rnage: 20 V@1 kHz) | 0.10 | mV/V |
| | | | 10 | V | 199 | V | ACV Source (Rnage: 200 V@1 kHz) | 0.10 | mV/V |
| | | | 100 | V | 1000 | V | ACV Source (Rnage: 1000 V@1 kHz) | 0.15 | mV/V |
| | | | 10 | mV | 10 | mV | ACV Meter (@1 Hz) | 1.3 | mV/V |
| | | | 20 | mV | 20 | mV | ACV Meter (@1 Hz) | 1.3 | mV/V |
| | | | 50 | mV | 50 | mV | ACV Meter (@1 Hz) | 1.3 | mV/V |
| | | | 100 | mV | 100 | mV | ACV Meter (@1 Hz) | 1.3 | mV/V |
| | | | 200 | mV | 200 | mV | ACV Meter (@1 Hz) | 0.64 | mV/V |
| | | | 500 | mV | 500 | mV | ACV Meter (@1 Hz) | 0.51 | mV/V |



| calibration items | working standard | calibration method | measurand level or range | | | | measurement conditions /independent variable | smallest uncertainty | |
|---|---|-----------------------------|--------------------------|---------------------|---------------|-------|--|----------------------|-------|
| | brand /model | document name /no. | minimum value | units | maximum value | units | explanation | value | units |
| KF1011 ACV Source ACV Meter (on-site calibration included) | Multimeter Fluke/8508A Calibrator Fluke/5700A Keysight /33622A | Document No.: B00-CD-214 | 1 | V | 1 | V | ACV Meter (@ 1 Hz) | 1.3 | mV/V |
| | | | 2 | V | 2 | V | ACV Meter (@ 1 Hz) | 1.3 | mV/V |
| | | | 5 | V | 5 | V | ACV Meter (@ 1 Hz) | 0.51 | mV/V |
| | | | 7 | V | 7 | V | ACV Meter (@ 1 Hz) | 0.55 | mV/V |
| | | | 10 | mV | 10 | mV | ACV Meter (@ 2 Hz) | 1.3 | mV/V |
| | | | 20 | mV | 20 | mV | ACV Meter (@ 2 Hz) | 1.3 | mV/V |
| | | | 50 | mV | 50 | mV | ACV Meter (@ 2 Hz) | 1.3 | mV/V |
| | | | 100 | mV | 100 | mV | ACV Meter (@ 2 Hz) | 1.3 | mV/V |
| | | | 200 | mV | 200 | mV | ACV Meter (@ 2 Hz) | 0.63 | mV/V |
| | | | 500 | mV | 500 | mV | ACV Meter (@ 2 Hz) | 0.51 | mV/V |
| | | | 1 | V | 1 | V | ACV Meter (@ 2 Hz) | 1.3 | mV/V |
| | | | 2 | V | 2 | V | ACV Meter (@ 2 Hz) | 1.3 | mV/V |
| | | | 5 | V | 5 | V | ACV Meter (@ 2 Hz) | 0.51 | mV/V |
| | | | 7 | V | 7 | V | ACV Meter (@ 2 Hz) | 0.54 | mV/V |
| | | | 10 | mV | 10 | mV | ACV Meter (@ 5 Hz) | 1.3 | mV/V |
| | | | 20 | mV | 20 | mV | ACV Meter (@ 5 Hz) | 1.3 | mV/V |
| | | | 50 | mV | 50 | mV | ACV Meter (@ 5 Hz) | 1.3 | mV/V |
| | | | 100 | mV | 100 | mV | ACV Meter (@ 5 Hz) | 1.3 | mV/V |
| | | | 200 | mV | 200 | mV | ACV Meter (@ 5 Hz) | 0.63 | mV/V |
| | | | 500 | mV | 500 | mV | ACV Meter (@ 5 Hz) | 0.51 | mV/V |
| 1 | V | 1 | V | ACV Meter (@ 5 Hz) | 1.3 | mV/V | | | |
| 2 | V | 2 | V | ACV Meter (@ 5 Hz) | 1.3 | mV/V | | | |
| 5 | V | 5 | V | ACV Meter (@ 5 Hz) | 0.51 | mV/V | | | |
| 7 | V | 7 | V | ACV Meter (@ 5 Hz) | 0.54 | mV/V | | | |
| 10 | mV | 10 | mV | ACV Meter (@ 10 Hz) | 1.3 | mV/V | | | |
| 20 | mV | 20 | mV | ACV Meter (@ 10 Hz) | 1.3 | mV/V | | | |



| calibration items | working standard | calibration method | measurand level or range | | | | measurement conditions /independent variable | smallest uncertainty | |
|---|---|-----------------------------|--------------------------|----------------------------------|---------------|-------|--|----------------------|-------|
| | brand /model | document name /no. | minimum value | units | maximum value | units | explanation | value | units |
| KF1011 ACV Source ACV Meter (on-site calibration included) | Multimeter Fluke/8508A Calibrator Fluke/5700A Keysight /33622A | Document No.: B00-CD-214 | 50 | mV | 50 | mV | ACV Meter (@ 10 Hz) | 1.3 | mV/V |
| | | | 100 | mV | 100 | mV | ACV Meter (@ 10 Hz) | 1.3 | mV/V |
| | | | 200 | mV | 200 | mV | ACV Meter (@ 10 Hz) | 0.63 | mV/V |
| | | | 500 | mV | 500 | mV | ACV Meter (@ 10 Hz) | 0.51 | mV/V |
| | | | 1 | V | 1 | V | ACV Meter (@ 10 Hz) | 1.3 | mV/V |
| | | | 2 | V | 2 | V | ACV Meter (@ 10 Hz) | 1.3 | mV/V |
| | | | 5 | V | 5 | V | ACV Meter (@ 10 Hz) | 0.51 | mV/V |
| | | | 7 | V | 7 | V | ACV Meter (@ 10 Hz) | 0.54 | mV/V |
| | | | 10 | mV | 10 | mV | ACV Meter (@ 10 Hz) | 1.3 | mV/V |
| | | | 20 | mV | 20 | mV | ACV Meter (@ 20 Hz) | 1.3 | mV/V |
| | | | 50 | mV | 50 | mV | ACV Meter (@ 20 Hz) | 1.3 | mV/V |
| | | | 100 | mV | 100 | mV | ACV Meter (@ 20 Hz) | 1.3 | mV/V |
| | | | 200 | mV | 200 | mV | ACV Meter (@ 20 Hz) | 0.63 | mV/V |
| | | | 500 | mV | 500 | mV | ACV Meter (@ 20 Hz) | 0.51 | mV/V |
| | | | 1 | V | 1 | V | ACV Meter (@ 20 Hz) | 1.3 | mV/V |
| | | | 2 | V | 2 | V | ACV Meter (@ 20 Hz) | 1.3 | mV/V |
| | | | 5 | V | 5 | V | ACV Meter (@ 20 Hz) | 0.51 | mV/V |
| | | | 7 | V | 7 | V | ACV Meter (@ 20 Hz) | 0.54 | mV/V |
| | | | 10 | mV | 199 | mV | ACV Meter (Rnage: 200 mV@20 Hz) | 0.80 | mV/V |
| | | | 0.1 | V | 1.99 | V | ACV Meter (Rnage: 2 V@20 Hz) | 0.24 | mV/V |
| 1 | V | 19.9 | V | ACV Meter (Rnage: 20 V@20 Hz) | 0.24 | mV/V | | | |
| 10 | mV | 199 | mV | ACV Meter (Rnage: 200 mV@100 Hz) | 0.64 | mV/V | | | |
| 0.1 | V | 1.99 | V | ACV Meter (Rnage: 2 V@100 Hz) | 0.087 | mV/V | | | |
| 1 | V | 19.9 | V | ACV Meter (Rnage: 20 V@100 Hz) | 0.035 | mV/V | | | |
| 10 | mV | 199 | mV | ACV Meter (Rnage: 20 V@3 kHz) | 0.64 | mV/V | | | |
| 0.1 | V | 1.99 | V | ACV Meter (Rnage: 20 V@3 kHz) | 0.086 | mV/V | | | |



| calibration items | working standard | calibration method | measurand level or range | | | | measurement conditions /independent variable | smallest uncertainty | |
|---|--|-----------------------------|--------------------------|----------------------------------|---------------|-------|--|----------------------|-------|
| | brand /model | document name /no. | minimum value | units | maximum value | units | explanation | value | units |
| KF1011 ACV Source ACV Meter (on-site calibration included) | Multimeter Fluke/8508A Calibrator Fluke/5700A Keysight /33622A | Document No.: B00-CD-214 | 1 | V | 19.9 | V | ACV Meter (Rnage: 20 V@3 kHz) | 0.034 | mV/V |
| | | | 10 | mV | 199 | mV | ACV Meter (Rnage: 20 V@5 kHz) | 0.64 | mV/V |
| | | | 0.1 | V | 1.99 | V | ACV Meter (Rnage: 20 V@5 kHz) | 0.086 | mV/V |
| | | | 1 | V | 19.9 | V | ACV Meter (Rnage: 20 V@5 kHz) | 0.034 | mV/V |
| | | | 10 | mV | 199 | mV | ACV Meter (Rnage: 20 V@10 kHz) | 0.64 | mV/V |
| | | | 0.1 | V | 1.99 | V | ACV Meter (Rnage: 20 V@10 kHz) | 0.086 | mV/V |
| | | | 1 | V | 19.9 | V | ACV Meter (Rnage: 20 V@10 kHz) | 0.034 | mV/V |
| | | | 1 | mV | 10 | mV | ACV Meter (Rnage: 200 mV@50 Hz) | 2.0 | mV/V |
| | | | 10 | mV | 199 | mV | ACV Meter (Rnage: 200 mV@50 Hz) | 0.42 | mV/V |
| | | | 0.1 | V | 1.9 | V | ACV Meter (Rnage: 2 V@50 Hz) | 0.088 | mV/V |
| | | | 1 | V | 19.9 | V | ACV Meter (Rnage: 20 V@50 Hz) | 0.058 | mV/V |
| | | | 10 | V | 199 | V | ACV Meter (Rnage: 200 V@50 Hz) | 0.063 | mV/V |
| | | | 100 | V | 1000 | V | ACV Meter (Rnage: 1000 V@50 Hz) | 0.10 | mV/V |
| | | | 1 | mV | 10 | mV | ACV Meter (Rnage: 200 mV@60 Hz) | 2.1 | mV/V |
| | | | 10 | mV | 199 | mV | ACV Meter (Rnage: 200 mV@60 Hz) | 0.40 | mV/V |
| | | | 0.1 | V | 1.9 | V | ACV Meter (Rnage: 2 V@60 Hz) | 0.088 | mV/V |
| | | | 1 | V | 19.9 | V | ACV Meter (Rnage: 20 V@60 Hz) | 0.058 | mV/V |
| | | | 10 | V | 199 | V | ACV Meter (Rnage: 200 V@60 Hz) | 0.063 | mV/V |
| | | | 100 | V | 1000 | V | ACV Meter (Rnage: 1000 V@60 Hz) | 0.10 | mV/V |
| | | | 1 | mV | 10 | mV | ACV Meter (Rnage: 200 mV@400 Hz) | 2.1 | mV/V |
| 10 | mV | 199 | mV | ACV Meter (Rnage: 200 mV@400 Hz) | 0.48 | mV/V | | | |
| 0.1 | V | 1.9 | V | ACV Meter (Rnage: 2 V@400 Hz) | 0.087 | mV/V | | | |
| 1 | V | 19.9 | V | ACV Meter (Rnage: 20 V@400 Hz) | 0.053 | mV/V | | | |
| 10 | V | 199 | V | ACV Meter (Rnage: 200 V@400 Hz) | 0.061 | mV/V | | | |
| 100 | V | 1000 | V | ACV Meter (Rnage: 1000 V@400 Hz) | 0.10 | mV/V | | | |
| 1 | mV | 10 | mV | ACV Meter (Rnage: 200 mV@1 kHz) | 2.1 | mV/V | | | |



| calibration items | working standard | calibration method | measurand level or range | | | | measurement conditions /independent variable | smallest uncertainty | |
|---|---|-----------------------------|--------------------------|-------|---------------|-------|--|----------------------|-------|
| | brand /model | document name /no. | minimum value | units | maximum value | units | explanation | value | units |
| KF1011 ACV Source ACV Meter (on-site calibration included) | Multimeter Fluke/8508A Calibrator Fluke/5700A Keysight /33622A | Document No.: B00-CD-214 | 10 | mV | 199 | mV | ACV Meter (Rnage: 200 mV@1 kHz) | 0.47 | mV/V |
| | | | 0.1 | V | 1.9 | V | ACV Meter (Rnage: 2 V@1 kHz) | 0.089 | mV/V |
| | | | 1 | V | 19.9 | V | ACV Meter (Rnage: 20 V@1 kHz) | 0.053 | mV/V |
| | | | 10 | V | 199 | V | ACV Meter (Rnage: 200 V@1 kHz) | 0.056 | mV/V |
| | | | 100 | V | 1000 | V | ACV Meter (Rnage: 1000 V@1 kHz) | 0.10 | mV/V |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |
| KF1012 ACA Source AC Ampere Meter (on-site calibration included) | Multimeter Fluke/8508A Calibrator Fluke/5700A | Document No.: B00-CD-212 | 10 | μA | 190 | μA | (Range: 200 μA@50 Hz, 60 Hz, 400 Hz, 1 kHz) | 2.1 | mA/A |
| | | | 0.1 | mA | 1.9 | mA | (Range: 2 mA@50 Hz, 60 Hz, 400 Hz, 1 kHz) | 0.53 | mA/A |
| | | | 1 | mA | 19 | mA | (Range: 20 mA@50 Hz, 60 Hz, 400 Hz, 1 kHz) | 0.46 | mA/A |
| | | | 10 | mA | 190 | mA | (Range: 200 mA@50 Hz, 60 Hz, 400 Hz, 1 kHz) | 0.45 | mA/A |
| | | | 0.1 | A | 1.9 | A | (Range: 2 A@50 Hz, 60 Hz, 400 Hz, 1 kHz) | 0.81 | mA/A |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |
| KF1013 High Digital Voltage Meter (on-site calibration included) High AC Power Source (on-site calibration included) | High Voltage Digital Meter KIKUSUI 149-10A | Document No.: B00-CD-407 | 1 | kV | 10 | kV | (@60 Hz) | 13 | mV/V |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |



| calibration items | working standard | calibration method | measurand level or range | | | | measurement conditions /independent variable | smallest uncertainty | |
|---|---|-----------------------------|--------------------------|-------|---------------|-------|--|----------------------|-------|
| | brand /model | document name /no. | minimum value | units | maximum value | units | explanation | value | units |
| KF1014 AC Current Meter (on-site calibration included) AC Current Source (on-site calibration included) | Calibrator/Current Amplifier Fluke/5700A Fluke/52120A | Document No.: B00-CD-397 | 2 | A | 120 | A | (@50 Hz, 60 Hz) | 0.69 | mA/A |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |
| KF1014 AC Clamp on Meter AC Current Meter AC Current Source (On-Site Calibration) | Current Amplifier Fluke/52120A Current coil Fluke/52120A /Coil 3KA Fluke/52120A /Coil 6KA | Document No.: B00-CD-465 | 0.1 | A | 2 | A | (@50 Hz) | 0.28 | mA/A |
| | | | 2 | A | 20 | A | (@50 Hz) | 0.21 | mA/A |
| | | | 20 | A | 120 | A | (@50 Hz) | 0.68 | mA/A |
| | | | 100 | A | 200 | A | (@50 Hz) | 0.78 | mA/A |
| | | | 200 | A | 1000 | A | (@50 Hz) | 0.87 | mA/A |
| | | | 1000 | A | 6000 | A | (@50 Hz) | 2.8 | mA/A |
| | | | 0.1 | A | 2 | A | (@60 Hz) | 0.25 | mA/A |
| | | | 2 | A | 20 | A | (@60 Hz) | 0.21 | mA/A |
| | | | 20 | A | 120 | A | (@60 Hz) | 0.68 | mA/A |
| | | | 100 | A | 200 | A | (@60 Hz) | 0.69 | mA/A |
| | | | 200 | A | 1000 | A | (@60 Hz) | 0.88 | mA/A |
| | | | 1000 | A | 6000 | A | (@60 Hz) | 2.8 | mA/A |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |



| calibration items | working standard | calibration method | measurand level or range | | | | measurement conditions /independent variable | smallest uncertainty | |
|--|---|-----------------------------|--------------------------|-------|---------------|-------|--|----------------------|-------|
| | brand /model | document name /no. | minimum value | units | maximum value | units | explanation | value | units |
| KF1099 Current Coil | Current Coil Fluke/52120A /COIL3KA Fluke/52120A /COIL6KA | Document No.: B00-CD-497 | 10 | | 10 | | DC | 0.60 | % |
| | | | 50 | | 50 | | DC | 0.60 | % |
| | | | 10 | | 10 | | 50 Hz | 0.47 | % |
| | | | 50 | | 50 | | 50 Hz | 0.47 | % |
| | | | 10 | | 10 | | 60 Hz | 0.47 | % |
| | | | 50 | | 50 | | 60 Hz | 0.47 | % |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |
| KF3001 Ground Bond Testers | Standard Resistor BURSTER/1282 | Document No.: B00-CD-404 | 10 | mΩ | 10 | mΩ | (@25 A, 60 Hz) | 59 | mΩ/Ω |
| | | | 10 | mΩ | 10 | mΩ | (@40 A, 60 Hz) | 59 | mΩ/Ω |
| | | | 10 | mΩ | 10 | mΩ | (@60 A, 60 Hz) | 59 | mΩ/Ω |
| | | | 50 | mΩ | 50 | mΩ | (@25 A, 60 Hz) | 16 | mΩ/Ω |
| | | | 50 | mΩ | 50 | mΩ | (@40 A, 60 Hz) | 16 | mΩ/Ω |
| | | | 50 | mΩ | 50 | mΩ | (@60 A, 60 Hz) | 16 | mΩ/Ω |
| | | | 100 | mΩ | 100 | mΩ | (@25 A, 60 Hz) | 13 | mΩ/Ω |
| | | | 100 | mΩ | 100 | mΩ | (@40 A, 60 Hz) | 13 | mΩ/Ω |
| | | | 100 | mΩ | 100 | mΩ | (@60 A, 60 Hz) | 13 | mΩ/Ω |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |
| KF3001 Resistor Ohmmeter (On-Side Calibration Included) | Multimeter Fluke/8508 Low Resistance Multiple Shunt OHM-LABS/7707 Current Transformer LEM/ITN1000-S | Document No.: B00-CD-567 | 5 | μΩ | 5 | μΩ | Standard Resistor | 6.5 | mΩ/Ω |
| | | | 10 | μΩ | 10 | μΩ | Standard Resistor | 6.5 | mΩ/Ω |
| | | | 20 | μΩ | 20 | μΩ | Standard Resistor | 6.5 | mΩ/Ω |
| | | | 50 | μΩ | 50 | μΩ | Standard Resistor | 6.5 | mΩ/Ω |
| | | | 100 | μΩ | 100 | μΩ | Standard Resistor | 6.5 | mΩ/Ω |
| | | | 200 | μΩ | 200 | μΩ | Standard Resistor | 6.5 | mΩ/Ω |
| | | | 500 | μΩ | 500 | μΩ | Standard Resistor | 6.5 | mΩ/Ω |



| calibration items | working standard | calibration method | measurand level or range | | | | measurement conditions /independent variable | smallest uncertainty | |
|--|---|--------------------------------------|--------------------------|-------------|---------------|-----------------------|--|----------------------|-----------------------|
| | brand /model | document name /no. | minimum value | units | maximum value | units | explanation | value | units |
| KF3001 Resistor Ohmmeter (On-Side Calibration Included) | Multimeter Fluke/8508 Low Resistance Multiple Shunt OHM-LABS/7707 Current Transformer LEM/ITN1000-S | Document No.: B00-CD-567 | 5 | $\mu\Omega$ | 5 | $\mu\Omega$ | Ohmmeter | 1.9 | m Ω / Ω |
| | | | 10 | $\mu\Omega$ | 10 | $\mu\Omega$ | Ohmmeter | 1.9 | m Ω / Ω |
| | | | 20 | $\mu\Omega$ | 20 | $\mu\Omega$ | Ohmmeter | 1.9 | m Ω / Ω |
| | | | 50 | $\mu\Omega$ | 50 | $\mu\Omega$ | Ohmmeter | 1.9 | m Ω / Ω |
| | | | 100 | $\mu\Omega$ | 100 | $\mu\Omega$ | Ohmmeter | 1.9 | m Ω / Ω |
| | | | 200 | $\mu\Omega$ | 200 | $\mu\Omega$ | Ohmmeter | 1.9 | m Ω / Ω |
| | | | 500 | $\mu\Omega$ | 500 | $\mu\Omega$ | Ohmmeter | 1.9 | m Ω / Ω |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |
| KF3001 Resistor Ohmmeter (On-Side Calibration Included) | Standard Resistor IET/HARS-X-8 IET/HRRS-B-6 FLUKE 8508A KEITHLEY 6517B YOKOGAWA 2792 | Document No.: B00-CD-442, B00-CD-443 | 0.001 | Ω | 0.001 | Ω | Resistor | 0.16 | m Ω / Ω |
| | | | 0.01 | Ω | 0.01 | Ω | Resistor | 0.22 | m Ω / Ω |
| | | | 0.02 | Ω | 0.02 | Ω | Resistor | 0.22 | m Ω / Ω |
| | | | 0.05 | Ω | 0.05 | Ω | Resistor | 0.20 | m Ω / Ω |
| | | | 0.1 | Ω | 0.1 | Ω | Resistor | 0.21 | m Ω / Ω |
| | | | 0.2 | Ω | 0.2 | Ω | Resistor | 0.034 | m Ω / Ω |
| | | | 0.5 | Ω | 0.5 | Ω | Resistor | 0.031 | m Ω / Ω |
| | | | 1.0 | Ω | 1.0 | Ω | Resistor | 0.031 | m Ω / Ω |
| | | | 2.0 | Ω | 2.0 | Ω | Resistor | 0.022 | m Ω / Ω |
| | | | 0.001 | Ω | 2.0 | Ω | Resistor | 0.22 | m Ω / Ω |
| | | | 2 | Ω | 20 | Ω | Resistor | 0.028 | m Ω / Ω |
| | | | 20 | Ω | 200 | Ω | Resistor | 0.021 | m Ω / Ω |
| | | | 200 | Ω | 2 | k Ω | Resistor | 0.022 | m Ω / Ω |
| | | | 2 | k Ω | 20 | k Ω | Resistor | 0.021 | m Ω / Ω |
| | | | 20 | k Ω | 200 | k Ω | Resistor | 0.026 | m Ω / Ω |
| 200 | k Ω | 2.0 | M Ω | Resistor | 0.10 | m Ω / Ω | | | |
| 2 | M Ω | 20 | M Ω | Resistor | 0.10 | m Ω / Ω | | | |



| calibration items | working standard | calibration method | measurand level or range | | | | measurement conditions /independent variable | smallest uncertainty | |
|---|--|--------------------------------------|--------------------------|----------|---------------|-------|--|----------------------|-------|
| | brand /model | document name /no. | minimum value | units | maximum value | units | explanation | value | units |
| KF3001 Resistor Ohmmeter (On-Side Calibration Included) | Standard Resistor IET/HARS-X-8 IET/HRRS-B-6 FLUKE 8508A KEITHLEY 6517B YOKOGAWA 2792 | Document No.: B00-CD-442, B00-CD-443 | 20 | MΩ | 200 | MΩ | Resistor | 0.32 | mΩ/Ω |
| | | | 200 | MΩ | 2 | GΩ | Resistor | 0.69 | mΩ/Ω |
| | | | 1 | GΩ | 10 | GΩ | Resistor | 3.7 | mΩ/Ω |
| | | | 10 | GΩ | 100 | GΩ | Resistor | 6.7 | mΩ/Ω |
| | | | 100 | GΩ | 10000 | GΩ | Resistor | 15 | mΩ/Ω |
| | | | 0.001 | Ω | 0.001 | Ω | Ohmmeter | 0.14 | mΩ/Ω |
| | | | 0.01 | Ω | 0.01 | Ω | Ohmmeter | 0.21 | mΩ/Ω |
| | | | 0.02 | Ω | 0.02 | Ω | Ohmmeter | 0.21 | mΩ/Ω |
| | | | 0.05 | Ω | 0.05 | Ω | Ohmmeter | 0.20 | mΩ/Ω |
| | | | 0.1 | Ω | 0.1 | Ω | Ohmmeter | 0.20 | mΩ/Ω |
| | | | 0.2 | Ω | 0.2 | Ω | Ohmmeter | 0.032 | mΩ/Ω |
| | | | 0.5 | Ω | 0.5 | Ω | Ohmmeter | 0.031 | mΩ/Ω |
| | | | 1 | Ω | 1 | Ω | Ohmmeter | 0.031 | mΩ/Ω |
| | | | 2 | Ω | 2 | Ω | Ohmmeter | 0.021 | mΩ/Ω |
| | | | 0.001 | Ω | 2 | Ω | Ohmmeter | 0.21 | mΩ/Ω |
| | | | 2 | Ω | 20 | Ω | Ohmmeter | 0.024 | mΩ/Ω |
| | | | 20 | Ω | 200 | Ω | Ohmmeter | 0.021 | mΩ/Ω |
| | | | 200 | Ω | 2 | kΩ | Ohmmeter | 0.021 | mΩ/Ω |
| | | | 2 | kΩ | 20 | kΩ | Ohmmeter | 0.021 | mΩ/Ω |
| | | | 20 | kΩ | 200 | kΩ | Ohmmeter | 0.026 | mΩ/Ω |
| 200 | kΩ | 2 | MΩ | Ohmmeter | 0.10 | mΩ/Ω | | | |
| 2 | MΩ | 20 | MΩ | Ohmmeter | 0.10 | mΩ/Ω | | | |
| 20 | MΩ | 200 | MΩ | Ohmmeter | 0.31 | mΩ/Ω | | | |
| 200 | MΩ | 2 | GΩ | Ohmmeter | 0.51 | mΩ/Ω | | | |
| 1 | GΩ | 10 | GΩ | Ohmmeter | 3.6 | mΩ/Ω | | | |



| calibration items | working standard | calibration method | measurand level or range | | | | measurement conditions /independent variable | smallest uncertainty | |
|--|--|--------------------------------------|--------------------------|-------------|---------------|-------|--|----------------------|-------|
| | brand /model | document name /no. | minimum value | units | maximum value | units | explanation | value | units |
| KF3001 Resistor Ohmmeter (On-Side Calibration Included) | Standard Resistor IET/HARS-X-8 | Document No.: B00-CD-442, B00-CD-443 | 10 | GΩ | 100 | GΩ | Ohmmeter | 5.4 | mΩ/Ω |
| | FLUKE 8508A KEITHLEY 6517B YOKOGAWA 2792 | | 100 | GΩ | 1000 | GΩ | Ohmmeter | 15 | mΩ/Ω |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |
| KF3001 Resistor Ohmmeter (on-site calibration included) | Multimeter/Calibrator /Resistance standard Fluke/8508A Fluke/5700A Wavetek/1271 Fluke/742A-1 | Document No.: B00-CD-215 | 1 | Ω | 1 | Ω | 4W Ohmmeter | 1.2 | μΩ/Ω |
| | | | 10 | Ω | 10 | Ω | 4W Ohmmeter | 7.3 | μΩ/Ω |
| | | | 100 | Ω | 100 | Ω | 4W Ohmmeter | 6.6 | μΩ/Ω |
| | | | 1 | kΩ | 1 | kΩ | 4W Ohmmeter | 3.3 | μΩ/Ω |
| | | | 10 | kΩ | 10 | kΩ | 4W Ohmmeter | 3.4 | μΩ/Ω |
| | | | 100 | kΩ | 100 | kΩ | 4W Ohmmeter | 12 | μΩ/Ω |
| | | | 1 | MΩ | 1 | MΩ | 4W Ohmmeter | 6.7 | μΩ/Ω |
| | | | 10 | MΩ | 10 | MΩ | 4W Ohmmeter | 18 | μΩ/Ω |
| | | | 100 | MΩ | 100 | MΩ | 2W Ohmmeter | 60 | μΩ/Ω |
| | | | 1 | Ω | 1 | Ω | 4W Resistor | 5.8 | μΩ/Ω |
| | | | 10 | Ω | 10 | Ω | 4W Resistor | 8.3 | μΩ/Ω |
| | | | 100 | Ω | 100 | Ω | 4W Resistor | 6.5 | μΩ/Ω |
| | | | 1 | kΩ | 1 | kΩ | 4W Resistor | 3.1 | μΩ/Ω |
| | | | 10 | kΩ | 10 | kΩ | 4W Resistor | 3.1 | μΩ/Ω |
| | | | 100 | kΩ | 100 | kΩ | 4W Resistor | 6.4 | μΩ/Ω |
| | | | 1 | MΩ | 1 | MΩ | 4W Resistor | 4.3 | μΩ/Ω |
| 10 | MΩ | 10 | MΩ | 4W Resistor | 16 | μΩ/Ω | | | |



| calibration items | working standard | calibration method | measurand level or range | | | | measurement conditions /independent variable | smallest uncertainty | |
|--|--|--------------------------|--------------------------|-------|---------------|-------|--|----------------------|-------|
| | brand /model | document name /no. | minimum value | units | maximum value | units | explanation | value | units |
| KF3001 Resistor Ohmmeter (on-site calibration included) | Multimeter/Calibrator /Resistance standard Fluke/8508A Fluke/5700A Wavetek/1271 Fluke/742A-1 | Document No.: B00-CD-215 | 100 | MΩ | 100 | MΩ | 2W Resistor | 56 | μΩ/Ω |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |



Time And Frequency

| calibration items | working standard | calibration method | measurand level or range | | | | measurement conditions /independent variable | smallest uncertainty | |
|--|--|--------------------------|--------------------------|-------------|---------------|-------------|--|-------------------------|-------|
| | brand /model | document name /no. | minimum value | units | maximum value | units | explanation | value | units |
| KJ0100 Stop Watch non-contact counter (on-site calibration included) | Rubidium Atomic Frequency Standard WAVETEK/909 Signal Generator Keysight/33622A | Document No.: B00-CD-417 | | | | | time base measurement (@32768 Hz) | 1.3x10 ⁻⁶ | |
| | | | 1 | s | 1 | h | direct comparison | 0.058 | s |
| | | | 1 | h | 20 | h | direct comparison | 0.097 | s |
| | | | 20 | h | 24 | h | direct comparison | 0.12 | s |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |
| KJ0200 Frequency counter / Frequency generator/ Function generator/ Spectrum analyzer/ Network analyzer (on-site calibration included) | PSG Analog Signal Generator & Frequency Doubler Multiplier AGILENT E8257D & Marki D1550MN | Document No.: B00-CD-471 | 10 (0.1) | MHz (μs) | 60 (16.7) | GHz (ps) | | 9.4 x 10 ⁻¹⁰ | |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |
| KJ0200 Frequency generator Signal Generator Spectrum Analyzer (on-site calibration included) | Rubidium Atomic Frequency Standard WAVETEK/909 Signal Generator Keysight/33622A | Document No.: B00-CD-424 | 0.001 (1000) | Hz (s) | 0.01 (100) | Hz (s) | | 2.7 x 10 ⁻⁸ | |
| | | | 0.01 (100) | Hz (s) | 1.0 (1) | Hz (s) | | 2.8 x 10 ⁻⁹ | |
| | | | 1.0 (1) | Hz (s) | 10 (0.1) | MHz (μs) | | 8.1 x 10 ⁻¹⁰ | |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |



| calibration items | working standard | calibration method | measurand level or range | | | | measurement conditions /independent variable | smallest uncertainty | |
|---|--|--------------------------|--------------------------|-------|---------------|-------|--|----------------------|-------|
| | brand /model | document name /no. | minimum value | units | maximum value | units | explanation | value | units |
| KJ0300 Dodging frequency meter Tachometer Centrifuge (on-site calibration included) | Rubidium Atomic Frequency Standard WAVETEK/909 Signal Generator Keysight/33622A | Document No.: B00-CD-425 | 0.6 | rpm | 6 | rpm | Dodging frequency meter | 0.00085 | rpm |
| | | | 6 | rpm | 150 | rpm | Dodging frequency meter | 0.0084 | rpm |
| | | | 150 | rpm | 1500 | rpm | Dodging frequency meter | 0.061 | rpm |
| | | | 1500 | rpm | 15000 | rpm | Dodging frequency meter | 0.61 | rpm |
| | | | 15000 | rpm | 150000 | rpm | Dodging frequency meter | 6.1 | rpm |
| | | | 150000 | rpm | 600000 | rpm | Dodging frequency meter | 24 | rpm |
| | | | 2 | rpm | 5 | rpm | Tachometer | 0.0071 | rpm |
| | | | 5 | rpm | 100 | rpm | Tachometer | 0.011 | rpm |
| | | | 100 | rpm | 1000 | rpm | Tachometer | 0.088 | rpm |
| | | | 1000 | rpm | 10000 | rpm | Tachometer | 0.89 | rpm |
| | | | 10000 | rpm | 99990 | rpm | Tachometer | 6.1 | rpm |
| | | | 200 | rpm | 12000 | rpm | Centrifuge | 45 | rpm |
| | | | 600000 | rpm | 1000000 | rpm | Dodging frequency meter | 40 | rpm |
| Approval Signatory: WANG, Vick; MOU, Wan-Chau; HSU, Chi-Shu; YEH, Ming-Hai; CHUNG, Po-Hsiung | | | | | | | | | |

Note: Smallest uncertainty represents an expanded uncertainty using a coverage factor approximately 95 % level of confidence.
(Null Below)

